Stores & Purchase Department

Rajiv Gandhi University of Knowledge Technologies-RK Valley

(A.P. Government ACT 18 of 2008)

Rajiv Knowledge Valley (Idupulapaya), Vempalli (M), Y.S.R. Kadapa Dist. AP-516330

Tel.No: 08588-283612 Email: purchasedept@rguktrkv.ac.in

NIQ. No: RGUKT-RKV/SPD/Chemistry/Lab Equipment/Q55C 01/12/2016

NOTICE INVITING QUOTATIONS

Sealed quotations are hereby invited from the registered firms/vendors for supply of following item to the RGUKT-RK Valley with term and conditions mentioned below.

S.NO	Item Description	Qty	Unit Price (Rs/-)	Taxes (%)	Total (Taxes& Transportation) Rs/-
1.	Digital Colorimeter as per Annexure-A	10 No's			

Rupees.....only

Terms & Conditions:

1. Sealed covers containing quotations must reach this office on or before **08.12.2016 by 05:00PM**, through Speed-post/registered Post/by hand.
2. The covers must be super scribed with "**Ref: RGUKT-RKV/SPD/Chemistry/Lab Equipment/Q55C dt 01.12.2016**".
3. The quotations must be addressed to "The Director, RGUKT, RK Valley, Idupulapaya"
4. The items must be delivered to RGUKT-RK Valley. The prices are to be quoted for destination.
5. VAT and other taxes of the above items should be mentioned clearly in the quotation against the item.
6. The items must be delivered within 05 days from the date of purchase order.
7. Payment shall be made 100% after delivery of the goods in good condition.

8. The items shall be delivered to RGUKT, RK Valley on door delivery basis. Unloading of items and delivery to the store place shall be responsibility of the firm.
9. RGUKT reserves the right to reject any/all quotation(s) without assigning any reasons whatsoever.
10. The quantity mentioned may change as per our requirements.

Address for sending of sealed quotations:

Stores & Purchase Department,
RGUKT- RK Valley (AP-IIIT),
Vemaplli Mandal,
YSR Kadapa Dist,
Andhrapradesh
Pin: 516330
Call: 09943713627, 9030257400

Sd/-

Director

Annexure-A

The following item is required for conducting regular lab experiments.

NAME OF THE ITEM	Digital Colorimeter
NO OF ITEM REQUIRED	10 No’s
LIGHT SOURCE	LED source
DETECTOR:	FOTO DIOD
FILTERS:	EIGHT (400-800nm)
DISPLAY:	RED LED Display -3 Digit
SOLUTION VOLUME:	1ml (min)
ACCURACY:	0.02 O.D
REPRODUCIBILITY:	0.01 O.D
STABILITY :	0.01 O.D